



OPEN ACCESS

PivNG primers and probes set used in the cobas omni Utility Channel is a reliable supplemental test for detection of *Neisseria gonorrhoeae* in oropharyngeal, urogenital and rectal specimens collected in cobas PCR Media

Mark Hopkins ^{1,2}, Rodney Arcenas,³ Xose Couto-Parada,² Michael Lewinski,³ Merlin Njoya,³ Deva Perinpanathan,² Rizwaan Sheriff,² Avneet Hansra,³ Selena Singh²

► Additional supplemental material is published online only. To view, please visit the journal online (<http://dx.doi.org/10.1136/sextrans-2022-055576>).

¹Department of Infection and Immunity, Liverpool University Hospitals NHS Foundation Trust, Liverpool, UK

²Department of Infection and Immunity, Barts Health NHS Trust, London, UK

³Roche Molecular Systems Inc, Pleasanton, California, USA

Correspondence to

Dr Mark Hopkins, Department of Infection and Immunity, Liverpool University Hospitals NHS Foundation Trust, Liverpool, UK; mark.hopkins@liverpoolft.nhs.uk

Received 14 July 2022

Accepted 1 April 2023

Published Online First

27 April 2023



© Author(s) (or their employer(s)) 2023. Re-use permitted under CC BY. Published by BMJ.

To cite: Hopkins M, Arcenas R, Couto-Parada X, et al. *Sex Transm Infect* 2023;**99**:416–419.

ABSTRACT

Objective To evaluate the clinical performance of the novel PivNG primers and probes set (PivNG test) used in the cobas omni Utility Channel for supplemental testing of *Neisseria gonorrhoeae* (NG).

Methods Oropharyngeal, urogenital and rectal samples were self-collected during routine testing at Barts Health sexual health clinics, London, UK. Samples were tested by the cobas CT/NG test and PivNG cobas omni Utility Channel test on cobas 6800/8800 Systems. Supplemental testing was carried out with the Xpert CT/NG test. PivNG overall percent agreements, positive percent agreements (PPAs)/negative percent agreements (NPAs) and positive/negative predictive values were calculated for each sample type. Microscopy and/or culture data were included for a randomised subset of concordant/discordant results, and a composite reference standard (cobas CT/NG, Xpert CT/NG and culture results) adjusted for partial verification bias was used to determine PivNG PPA and NPA.

Results Of 447 evaluable samples with valid results from all three assays (cobas CT/NG, PivNG and Xpert CT/NG), 288 (64.4%) were NG-positive by both PivNG and cobas CT/NG; 117 (26.2%) were NG-negative in both tests; and 42 (9.4%) had discordant results (with NG-negative supplementary Xpert) CT/NG results in 40/42 instances). Of 19 PivNG/Xpert CT/NG-discordant samples, 11 were confirmed NG-positive by microscopy and/or culture results. PivNG PPA and NPA were 100% and 91% for oropharyngeal swabs, 100% and 100% for vaginal swabs, 100% and 100% for male urine samples, and 100% and 97% for rectal swabs, respectively, compared with the partially adjusted composite reference standard.

Conclusions PivNG is a reliable supplementary test with high sensitivity for confirming NG infection when used in conjunction with the cobas CT/NG test and samples collected in cobas PCR Media. Moreover, the PivNG test offers a convenient, high-throughput solution for supplemental NG testing of various sample types, with the potential to reduce the number of indeterminate reports.

INTRODUCTION

Neisseria gonorrhoeae (NG) is the second most common bacterial STI,^{1,2} causing an estimated 82 million new cases of gonorrhoea worldwide in 2020.¹ Transmission is by direct mucosal contact and

can lead to symptomatic or asymptomatic infections in the urethra, endocervix, rectum and pharynx.²

NG is diagnosed by detection using a sample collected from the infected site with nucleic acid amplification tests (NAATs), culture or direct microscopy.² NAATs are more sensitive and easier to perform and provide results faster than culture;^{2,3} however, cross-reactivity can occur due to commensal *Neisseria* species at the same sites.³ Therefore, UK screening guidelines recommend a supplementary NAAT if the positive predictive value (PPV) of the initial test is <90% for the population tested.^{3,4}

The cobas CT/NG test for use on cobas 6800/8800 Systems (Roche Molecular Systems Inc., USA) is a NAAT for direct detection of *Chlamydia trachomatis* (CT) and/or NG DNA from various specimens, including extragenital sources.⁵ The cobas CT/NG test targets a highly conserved direct repeat (DR9) region of the NG genome.

The cobas 6800/8800 Systems are fully automated and high throughput, designed primarily for running commercial/CE-marked PCR assays, and they have open-channel functionality capable of performing laboratory-developed tests on the cobas omni Utility Channel. A novel alternative primers and probes set targeting the NG-specific *pilin inversion* (*piv*) gene was provided by Integrated DNA Technologies for use on this channel.⁶ Using the cobas omni Utility Channel for supplementary testing removes the need for additional laboratory equipment.

The objective of this study was to evaluate the clinical performance of the PivNG primers and probes set (PivNG test) for supplemental testing of NG in oropharyngeal, urogenital and rectal samples.

METHODS

Study design and specimen collection

Oropharyngeal, vaginal, male urine and rectal specimens were self-collected in cobas PCR Media by patients attending routine testing for STIs at Barts Health sexual health clinics, London, UK. Patients unwilling or unable to self-collect samples were excluded. Samples were obtained and tested from November 2019 to June 2020. Samples were initially tested using the cobas CT/NG test as part of routine screening. Only samples with sufficient

surplus volume for PivNG and Xpert CT/NG test (Cepheid) supplemental testing were included in the evaluation.

Sample testing

Specimens were tested for NG with the CE-IVD cobas CT/NG test and also the PivNG cobas omni Utility Channel test on the cobas 6800/8800 Systems, following the manufacturer's recommended instructions for use. Supplemental testing was also carried out using the Xpert CT/NG test by adding 300 µL of sample collected in cobas PCR Media to the Cepheid cartridge. NG culture and microscopy results were used, if available, for concordant/discordant sample analysis.

Although all samples were self-collected by the patient, neither the cobas CT/NG test nor the Xpert CT/NG test is licensed for use with patient self-collected oral or anorectal swabs. Self-collected vaginal swabs are on-label for both assays. Self-collected extra-genital swabs were validated locally by BARTS Health NHS Trust where samples were co-collected in Aptima and cobas media to compare performance with the Hologic Panther.

Data analysis

All analyses were performed using SAS/STAT software V.9.4. The overall percent agreement (OPA), positive percent agreement (PPA), negative percent agreement (NPA), PPV and negative predictive value of the PivNG test with respect to specimen type and compared with the cobas CT/NG test were calculated, along with corresponding two-sided 95% CIs using Wilson's score method. A sample size of 60 ensured a precision level (half-width of 95% CI) of 3%–8%, assuming agreement of 90–100%. Discordant analysis evaluated results of NAATs, microscopy and culture, as explained in the online supplemental methods.

Invalid/failed tests were excluded from the data analyses. Analytical sensitivities of the PivNG and Xpert CT/NG tests were determined, as explained in the online supplemental methods. Analytical specificity of the PivNG test was determined using a panel of 20 non-gonococcal isolates. NPA was determined by (number of PivNG-negative results/total number of non-gonococcal isolates)×100, and CIs were calculated using the Clopper-Pearson exact binomial CI method.

RESULTS

Characteristics of clinical samples

Overall, 447 evaluable samples were obtained from 377 patients (28.9% (109) female and 70.3% (265) male, 0.8% (3) unknown), with a mean age of 32.5±10.8 years. Of the 377 patients who contributed evaluable samples, 83.6% (315) provided one type; 14.3% (54) provided two types; and 2.1% (8) provided three types.

Performance of the PivNG test for secondary detection of NG infection

The OPA was 90.6% between PivNG and cobas CT/NG tests, with full concordance for 288 of 447 (64.4%) samples, and 117 of 447 (26.2%) samples were NG-negative with both tests (table 1). Discordant PivNG and cobas CT/NG results were recorded for 42 (9.4%) samples, of which 40 were NG-negative using the supplementary Xpert CT/NG test. The majority (27 of 40) were oropharyngeal swabs (table 1). Culture and microscopy results were unavailable from this collection site. Microscopy and/or culture NG results were included for a randomised subset of concordant/discordant NAAT results (online supplemental table S1). A composite reference standard was created

Table 1 Diagnostic performance for detection of NG DNA using PivNG compared with the cobas CT/NG test and with reactive samples retested using Xpert CT/NG as a supplementary NAAT

	Positive or negative samples by assay (n)				Performance		
	cobas NG (+) and PivNG (+)	cobas NG (+) and PivNG (-)	cobas NG (-) and PivNG (+)	cobas NG (-) and PivNG (-)	OPA % (95% CI)	PPA% (95% CI), NPA % (95% CI)	PPV % (95% CI), NPV % (95% CI)
All samples*	288	40	2	117	90.6 (87.5 to 93.0)	87.8 (83.8 to 90.9), 98.3 (94.1 to 99.5)	99.3 (97.5 to 99.8), 74.5 (67.2 to 80.7)
Of which Xpert NG+	269	0	0				
Xpert NG-	19	40	2				
Oropharyngeal swab samples	87	27	2	38	81.2 (74.3 to 86.6)	76.3 (67.7 to 83.2), 95.0 (83.5 to 98.6)	97.8 (92.2 to 99.4), 58.5 (46.3 to 69.6)
Of which Xpert NG+	74	0	0				
Xpert NG-	13	27	2				
Vaginal swab samples	62	7	0	28	92.8 (85.9 to 96.5)	89.9 (80.5 to 95.0), 100 (87.9 to 100.0)	100 (94.2 to 100.0), 80.0 (64.1 to 90.0)
Of which Xpert NG+	60	0	0				
Xpert NG-	2	7	0				
Male urine samples	57	2	0	25	97.6 (91.7 to 99.3)	96.6 (88.5 to 99.1), 100 (86.7 to 100.0)	100 (93.7 to 100.0), 92.6 (76.6 to 97.9)
Of which Xpert NG+	56	0	0				
Xpert NG-	1	2	0				
Rectal swab samples	82	4	0	26	96.4 (91.2 to 98.6)	95.3 (88.6 to 98.2), 100 (87.1 to 100.0)	100 (95.5 to 100.0), 86.7 (70.3 to 94.7)
Of which Xpert NG +	79	0	0				
Xpert NG-	3	4	0				

Main analysis (cobas NG vs PivNG) values shown in bold

*Some patients provided more than one sample type.

+, positive; -, negative; CT, *Chlamydia trachomatis*; NAAT, nucleic acid amplification test; NG, *Neisseria gonorrhoeae*; NPA, negative percent agreement; NPV, negative predicted value; OPA, overall percent agreement; Piv, pilin inversion; PPA, positive percent agreement; PPV, positive predicted value.

using the cobas CT/NG, Xpert CT/NG and culture results. When adjusted for partial verification bias and compared with the composite reference standard, PivNG gave a PPA/NPA of 100%/91% for oropharyngeal swabs, 100%/100% for vaginal swabs, 100%/100% for male urine samples and 100%/97% for rectal swabs, respectively (online supplemental table S2).

When determining analytical performance, PivNG had a lower limit of detection for cobas NG-positive samples compared with Xpert CT/NG (online supplemental table S3) but retained high specificity when tested with non-gonococcal isolates (online supplemental table S4).

Supplementary Xpert CT/NG testing of samples in cobas PCR Media

NG-negative Xpert CT/NG results were also recorded for 19 samples with concordant NG-positive cobas CT/NG/PivNG results (13 oropharyngeal, 2 vaginal, 1 male urine and 3 rectal samples) (table 1). Oropharyngeal sites were not sampled for culture; however, urethral (n=10) or rectal (n=1) NG infection was confirmed by culture and/or microscopy in 11 of 19 of these patients (data not shown), indicating Xpert CT/NG false negativity. However, overall concordance between PivNG and Xpert CT/NG as secondary tests was good (40 negative and 269 positive) (table 1).

AmpliRun Total CT/NG/TV/MGE control (Viracell) material diluted in cobas PCR Media and molecular grade water indicated that cobas PCR Media was not inhibitory to Xpert CT/NG, but cobas CT/NG and PivNG reported positive results at lower concentrations using this protocol (online supplemental table S5).

DISCUSSION

Previous studies have established the performance of the cobas CT/NG test for detecting NG infection.^{5,7} This study evaluated clinical performance of the PivNG test as a supplementary test using the cobas 6800/8800 Systems in conjunction with the cobas CT/NG test. The PivNG test performed well with all specimen types and appeared a reliable secondary test for confirming NG infection using samples collected in cobas PCR Media. Improved sensitivity of PivNG compared with the Xpert CT/NG test in this setting was also supported by direct comparison of analytical performance.

Lower agreement between cobas CT/NG and PivNG tests was observed with oropharyngeal swabs compared with other sample types. Unfortunately, culture and microscopy results were unavailable for this subset of samples in our discordant analysis due to local practice, where the decision to sample different anatomical sites for culture/microscopy is guided by risk and symptoms at those sites during clinical presentation. Perry *et al* reported a lower PPV for detecting NG in oropharyngeal samples (88.6%) than rectal (96.4%) or urogenital (96.0%) samples.⁸ These data, along with the data from our study, highlight that rectal and urogenital samples may not require supplemental testing, but oropharyngeal samples should undergo secondary testing, in line with the UK guidelines.^{3,4} This is consistent with the literature regarding the presence of commensal *Neisseria*, particularly in the pharynx,³ and genetic exchange of NG DNA to these species, which can result in false positivity due to cross-reactivity with non-NG species when performing NAATs.⁹ While our results suggest good performance of the cobas CT/NG test and PivNG test with specimens collected in cobas PCR Media, we suggest that cobas NG-positive but secondary PivNG-negative samples are reported as equivocal or indeterminate, with

a repeat specimen request and subsequent PPV audit to inform a local reporting algorithm. The concern over cross-reactivity in extragenital samples needs to be balanced with the NAAT cycle threshold value, as samples with high values have an increased likelihood of non-confirmable results.¹⁰ Using highly sensitive screening, as well as reporting discordant samples as equivocal or indeterminate, allows therapy decisions to be influenced by pretest probability.

Our study has some limitations. The sample size was relatively small and varied across sample types. Determining PivNG performance in a larger sample set, with similar sized groups, would be beneficial to confirm results and enable clear comparisons across sample types. Analytical performance was evaluated only for the PivNG and Xpert CT/NG tests, not the cobas CT/NG test; the latter may have provided more clarity when evaluating discordant samples. Finally, all samples were routinely collected in cobas PCR Media, which is an off-label use for the Xpert CT/NG test and could affect performance. Brief evaluation of Xpert CT/NG analytical sensitivity with this protocol showed no evidence of inhibition, but the limit of detection appeared suboptimal. The collection of specimens in a single edia type is a real-world issue for laboratories when selecting supplementary tests. Our data show supplementary assay choice could influence the number of equivocal reports issued in settings where cobas PCR Media is used.

In conclusion, this is the first study to show the clinical performance of the PivNG test on the cobas omni Utility Channel. The performance of PivNG was good in conjunction with the cobas CT/NG test and could result in fewer equivocal reports being issued to clinic. This open-channel assay offers a convenient and high-throughput solution for supplemental NG testing of various sample types.

Correction notice This article has been corrected since it was first published online. The following paragraph has been added under the 'Sample testing' heading: Although all samples were self-collected by the patient, neither the cobas CT/NG test nor the Xpert CT/NG test is licensed for use with patient self-collected oral or anorectal swabs. Self-collected vaginal swabs are on-label for both assays. Self-collected extra-genital swabs were validated locally by BARTS Health NHS Trust where samples were co-collected in Aptima and cobas media to compare performance with the Hologic Panther.

Handling editor Apostolos Beloukas

Acknowledgements We thank Desiree Barcenilla at Barts Health NHS Trust for their laboratory assistance during the study. Medical writing support was provided by Katie Farrant at Elements Communications Ltd, Westerham, UK and was funded by Roche Diagnostics. The PivNG test is used for research purposes only and is not approved for clinical use. COBAS and COBAS OMNI are trademarks of Roche. All other product names and trademarks are the property of their respective owners.

Contributors MH, RA, XCP, ML and DP participated in the conceptualisation and design of the study. XCP, DP and RS were involved in the data collection. MH, XCP, ML, MN, DP, RS, AH and SS performed data analysis and interpretation. MH, RA, ML, MN, RS and AH contributed to the drafting of the manuscript. MH, RA, XCP, ML, MN, DP and AH and SS contributed to critical revision of the manuscript. All authors approved the final version of the manuscript.

Funding This study was funded by Roche Molecular Systems (study number: RD005106).

Competing interests MH received support from Roche Molecular Systems for study protocol costs and is a member of the speaker's bureau of Roche Molecular Systems. XCP received support from Roche Diagnostics to attend past meetings not related to this study. DP, RS and SS have nothing to disclose. RA, ML and AH are employees of Roche Molecular Systems and report receiving stocks and stock options from Roche. MN is an employee of Roche Molecular Systems.

Patient consent for publication Not applicable.

Ethics approval This study involves human participants but National Research Ethics Service approval was not required. This project was reviewed by Queen Mary University of London Joint Research Management Office and registered with Barts Health Clinical Effectiveness Unit (project ID 10468). Institutional review classified

the project as a clinical improvement project and not research. Ethical approval was not required because it was not research but quality assessment of a diagnostic assay for which the sample was originally obtained. Patient consent was obtained to test for *Chlamydia trachomatis* and *Neisseria gonorrhoeae*. The study was conducted in compliance with the International Conference on Harmonization Good Clinical Practice Guideline and local regulations.

Provenance and peer review Not commissioned; externally peer reviewed.

Supplemental material This content has been supplied by the author(s). It has not been vetted by BMJ Publishing Group Limited (BMJ) and may not have been peer-reviewed. Any opinions or recommendations discussed are solely those of the author(s) and are not endorsed by BMJ. BMJ disclaims all liability and responsibility arising from any reliance placed on the content. Where the content includes any translated material, BMJ does not warrant the accuracy and reliability of the translations (including but not limited to local regulations, clinical guidelines, terminology, drug names and drug dosages), and is not responsible for any error and/or omissions arising from translation and adaptation or otherwise.

Open access This is an open access article distributed in accordance with the Creative Commons Attribution 4.0 Unported (CC BY 4.0) license, which permits others to copy, redistribute, remix, transform and build upon this work for any purpose, provided the original work is properly cited, a link to the licence is given, and indication of whether changes were made. See: <https://creativecommons.org/licenses/by/4.0/>.

ORCID iD

Mark Hopkins <http://orcid.org/0000-0003-4744-8542>

REFERENCES

- 1 WHO. Sexually transmitted infections (STIs). 2021. Available: <https://www.who.int/news-room/fact-sheets/detail/sexually-transmitted-infections-stis>
- 2 Meyer T, Buder S. The laboratory diagnosis of *Neisseria gonorrhoeae*: current testing and future demands. *Pathogens* 2020;9:91.
- 3 Fifer H, Saunders J, Soni S, et al. 2018 UK national guideline for the management of infection with *Neisseria gonorrhoeae*. *Int J STD AIDS* 2020;31:4–15.
- 4 England Public Health. Guidance for the detection of gonorrhoea in England. 2021. Available: https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/972388/Guidance_for_the_detection_of_gonorrhoea_in_England_2021.pdf
- 5 Marlowe EM, Hardy D, Krevolin M, et al. High-throughput testing of urogenital and extragenital specimens for detection of *Chlamydia trachomatis* and *Neisseria gonorrhoeae* with cobas® CT/NG. *Eur J Microbiol Immunol (Bp)* 2017;7:176–86.
- 6 Integrated DNA Technologies. Microbial identification assays. 2022. Available: <https://eu.idtdna.com/pages/products/reagents-and-kits/microbial-detection/microbial-identification-assays>
- 7 Van Der Pol B, Fife K, Taylor SN, et al. Evaluation of the performance of the cobas CT/NG test for use on the COBAS 6800/8800 systems for detection of *Chlamydia trachomatis* and *Neisseria gonorrhoeae* in male and female urogenital samples. *J Clin Microbiol* 2019;57:e01996-18.
- 8 Perry MD, Jones RN, Corden SA. Is confirmatory testing of Roche cobas 4800 CT/NG test *Neisseria gonorrhoeae* positive samples required? comparison of the Roche cobas 4800 CT/NG test with an opa/pap duplex assay for the detection of *N gonorrhoeae*. *Sex Transm Infect* 2014;90:303–8.
- 9 Whiley DM, Tapsall JW, Sloots TP. Nucleic acid amplification testing for *Neisseria gonorrhoeae*: an ongoing challenge. *J Mol Diagn* 2006;8:3–15.
- 10 Bromhead C, Liyanarachchy N, Mayes J, et al. Supplementary testing is not required in the cobas 4800 CT/NG test for *Neisseria gonorrhoeae* weak-positive urogenital samples. *J Clin Microbiol* 2015;53:327–8.

Supplementary Material

The PivNG primers and probes set used in the cobas omni Utility Channel is a reliable supplemental test for detection of *Neisseria gonorrhoeae* in oropharyngeal, urogenital and rectal specimens collected in cobas PCR Media

Mark Hopkins, Rodney Arcenas, Xose Couto-Parada, Michael Lewinski, Merlin Njoya, Deva Perinpanathan, Rizwaan Sheriff, Avneet Hansra, Selena Singh

Supplementary Methods

Composite reference standard analysis of subset of discordant and concordant NAAT NG results with available microscopy and/or culture results

Given that microscopy and/or culture data were included for a randomised subset of samples with cobas[®] CT/NG and Xpert[®] CT/NG concordant and discordant NG results, the composite reference standard was partially verified. The randomised selection contained a total of 111 samples covering 59 discordant samples and a subset of concordant positive (36/269), concordant negative (5/40) and missing cobas/invalid Xpert (11/85) results. Of these, culture/microscopy results were available for 8 rectal (2 negative, 6 positive), 8 oropharyngeal (6 negative, 2 positive), 12 male urine (1 negative, 11 positive) and 6 vaginal (4 negative, 2 positive) samples. The Begg and Greenes' correction method¹ was used to calculate partial verification bias-adjusted estimates of PPA and NPA on the entire dataset (453 samples with valid PivNG results), with the assumption that the undetermined composite reference standard status was at random. Of this dataset, 6 samples were not evaluable (2 were missing both cobas CT/NG and Xpert CT/NG results and 4 had invalid Xpert CT/NG results) when determining the performance of PivNG compared with cobas CT/NG and Xpert CT/NG (Table 1; 447 samples).

Analytical performance

Analytical sensitivity, defined as the lowest concentration where $\geq 95\%$ of samples tested give positive results, was calculated using a probit curve fit to the lowest panel members. A panel per sample type with 25 replicates per level were tested in two specimen matrices over 5 days. Levels evaluated were 0.01, 0.03, 0.1, 0.3, 1 and 3 colony-forming units (CFUs)/mL for the swab matrix, and 0.005, 0.015, 0.05, 0.15, 0.5 and 1.5 CFUs/mL for the

urine matrix. Swab samples were collected in cobas PCR Media spiked with human colorectal carcinoma HCT-15 cells, and urine samples were collected in upper urinary tract matrix.

Limit of detection analysis of assays using a third-party control diluted in cobas PCR Media or molecular grade water

Amplirun® Total CT/NG/TV/MGE control material (Vircell) was serially diluted using either cobas PCR Media or molecular grade water to six dilution points with nominal concentrations of 111, 36, 8, 4, 1, 0.1 copies/mL in each media, spanning the limit of detection of the Xpert CT/NG test. The panel of 12 samples (six concentrations in each media) were tested in triplicate using cobas CT/NG, PivNG and Xpert CT/NG assays and results compared.

Reference

1. Arifin WN, Yusof UK. Correcting for partial verification bias in diagnostic accuracy studies: A tutorial using R. *Stat Med* 2022;41:1709–27.

Supplementary Table S1. Frequency of available microscopy and culture NG results, with PivNG results stratified by concordant and discordant NG results from cobas CT/NG and Xpert CT/NG

Specimen site	cobas/Xpert NG results	Microscopy and culture NG result	PivNG result	Frequency
Oropharyngeal swab	Concordant negative	Not available	NG negative	38
	Concordant negative	Not available	NG positive	2
	Concordant positive	Negative	NG positive	1
	Concordant positive	Not available	NG positive	72
	Concordant positive	Positive	NG positive	1
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	4
	Discordant positive/negative	Not available	NG negative	26
	Discordant positive/negative	Not available	NG positive	8
	Discordant positive/negative	Positive	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	2
Vaginal swab	Concordant positive	Negative	NG positive	2
	Concordant positive	Not available	NG positive	56
	Concordant positive	Positive	NG positive	2
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	1
	Discordant positive/negative	Not available	NG negative	6
	Discordant positive/negative	Not available	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	31
Male urine sample	Concordant positive	Not available	NG positive	45
	Concordant positive	Positive	NG positive	11
	Discordant positive/negative	Negative	NG negative	1

	Discordant positive/negative	Not available	NG negative	1
	Discordant positive/negative	Not available	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	25
Rectal swab	Concordant positive	Not available	NG positive	73
	Concordant positive	Positive	NG positive	6
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	1
	Discordant positive/negative	Not available	NG negative	3
	Discordant positive/negative	Not available	NG positive	2
	Missing cobas/invalid Xpert	Not available	NG negative	26
	Missing cobas/invalid Xpert	Not available	NG positive	1

CT, *Chlamydia trachomatis*; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion

Supplementary Table S2. PivNG PPA and NPA with a composite reference standard adjusted for partial verification bias using Begg and Greenes' method

Specimen site	PivNG	Composite reference standard (cobas and Xpert and culture)				PPA (95% CI)	NPA (95% CI)
		Positive	Negative	Undetermined	Total		
Oropharyngeal swab	Positive	75	6	8	89	100% (99.96, 100.00)	91.04% (84.25, 97.83)
	Negative	0	39	28	67		
	Total	75	45	36	156		
Vaginal swab	Positive	60	1	1	62	100% (99.07, 100.00)	100% (99.93, 100.00)
	Negative	0	1	37	38		
	Total	60	2	38	100		
Male urine sample	Positive	56	0	1	57	100% (99.07, 100.00)	100% (99.93, 100.00)
	Negative	0	1	26	27		
	Total	56	1	27	84		
Rectal swab	Positive	79	1	3	83	100% (99.28, 100.00)	96.66% (90.22, 100.00)
	Negative	0	1	29	30		
	Total	79	2	32	113		

CI, confidence interval; NG, *Neisseria gonorrhoeae*; NPA, negative percent agreement; Piv, pilin inversion; PPA, positive percent agreement

Supplementary Table S3. Summary limit of detection estimates for cobas NG-positive samples using PivNG compared with Xpert CT/NG

Matrix	Assay	LOD probit estimate (CFU/mL)	95% CI
Swab	PivNG	0.043	0.026, 0.199
	Xpert NG	0.326	0.162, 1.671
Urine	PivNG	0.018	0.011, 0.121
	Xpert NG	0.451	0.186, 3.136

CFU, colony-forming unit; CI, confidence interval; LOD, limit of detection; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion

Supplementary Table S4. NPA between a specificity panel of non-gonococcal isolates and PivNG-negative results

Expected result	Number of non-gonococcal isolates*	PivNG (+) results	PivNG (-) results	NPA (%)	95% exact CI
Negative	20	0	20	100	83.2, 100.0

*Non-gonococcal *Neisseria* isolates evaluated were: *N. cinerea*, *N. dentrificans*, *N. elongata subsp. elongata*, *N. elongata subsp. niroreducans*, *N. flava*, *N. flavescens*, *N. subflava*, *N. lactamica*, *N. macacae*, *N. meningitidis Serogroup A*, *N. meningitidis Serogroup B*, *N. meningitidis Serogroup C*, *N. meningitidis Serogroup D*, *N. meningitidis Serogroup W135*, *N. meningitidis Serogroup Y*, *N. mucosa*, *N. perflava*, *N. polysaccharea*, *N. sicca* and *N. subflava*

+, positive; -, negative; CI, confidence interval; NG, *Neisseria gonorrhoeae*; NPA, negative percent agreement; Piv, pilin inversion

Supplementary Table S5. Performance of Xpert CT/NG with Vircell control material diluted to six nominal concentrations using cobas PCR Media or molecular grade water

Vircell control diluted in cobas PCR Media					Vircell control diluted in molecular grade water				
NG copies/mL	Xpert NG*		cobas NG	PivNG	NG copies/mL	Xpert NG		cobas NG	PivNG
	NG2 Ct	NG4 Ct	NG Ct	PivNG Ct		NG2 Ct	NG4 Ct	NG Ct	PivNG Ct
111	32.4	32.7	34.22	30.4	111	31.9	31.8	34.14	30.6
111	32.5	32.6	34.28	30.5	111	31.9	31.5	33.84	30.5
111	32.6	32.2	33.84	30.3	111	32.3	32.1	34.75	30.6
36	35	34.8	36.53	32.2	36	34.6	33.5	35.67	32.2
36	33.9	34.1	35.92	31.8	36	34.4	33.7	36.71	32
36	34.4	33.5	36.02	32.4	36	35	34.4	35.27	32.3
8	Not detected	36.1	36.69	35	8	37.4	Not detected	39.08	34.3
8	Not detected	37.7	37.5	34.5	8	35.7	39.4	37.8	34.3
8	37.2	37.2	39.62	34.3	8	Not detected	37.1	38.28	35
4	Not detected	37.4	40.06	35.7	4	36.1	37.9	39.27	35.1
4	37.7	37.4	38.8	36.5	4	Not detected	37.9	41.67	35.2
4	36.7	Not detected	40.7	35.3	4	Not detected	35.9	Not detected	35.5
1	Not detected	Not detected	Not detected	Not detected	1	Not detected	Not detected	39.15	Not detected
1	Not detected	Not detected	39.83	35.8	1	Not detected	Not detected	Not detected	37.4
1	Not detected	Not detected	Not detected	Not detected	1	Not detected	Not detected	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	42	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	Not detected	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	Not detected	Not detected	Not detected

*Xpert CT/NG requires both NG targets (designated by Cepheid as NG2 and NG4) to be detected to assign a positive result

CT, *Chlamydia trachomatis*; Ct, cycle threshold; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion

Supplementary Material

The PivNG primers and probes set used in the cobas omni Utility Channel is a reliable supplemental test for detection of *Neisseria gonorrhoeae* in oropharyngeal, urogenital and rectal specimens collected in cobas PCR Media

Mark Hopkins, Rodney Arcenas, Xose Couto-Parada, Michael Lewinski, Merlin Njoya, Deva Perinpanathan, Rizwaan Sheriff, Avneet Hansra, Selena Singh

Supplementary Methods

Composite reference standard analysis of subset of discordant and concordant NAAT NG results with available microscopy and/or culture results

Given that microscopy and/or culture data were included for a randomised subset of samples with cobas[®] CT/NG and Xpert[®] CT/NG concordant and discordant NG results, the composite reference standard was partially verified. The randomised selection contained a total of 111 samples covering 59 discordant samples and a subset of concordant positive (36/269), concordant negative (5/40) and missing cobas/invalid Xpert (11/85) results. Of these, culture/microscopy results were available for 8 rectal (2 negative, 6 positive), 8 oropharyngeal (6 negative, 2 positive), 12 male urine (1 negative, 11 positive) and 6 vaginal (4 negative, 2 positive) samples. The Begg and Greenes' correction method¹ was used to calculate partial verification bias-adjusted estimates of PPA and NPA on the entire dataset (453 samples with valid PivNG results), with the assumption that the undetermined composite reference standard status was at random. Of this dataset, 6 samples were not evaluable (2 were missing both cobas CT/NG and Xpert CT/NG results and 4 had invalid Xpert CT/NG results) when determining the performance of PivNG compared with cobas CT/NG and Xpert CT/NG (Table 1; 447 samples).

Analytical performance

Analytical sensitivity, defined as the lowest concentration where $\geq 95\%$ of samples tested give positive results, was calculated using a probit curve fit to the lowest panel members. A panel per sample type with 25 replicates per level were tested in two specimen matrices over 5 days. Levels evaluated were 0.01, 0.03, 0.1, 0.3, 1 and 3 colony-forming units (CFUs)/mL for the swab matrix, and 0.005, 0.015, 0.05, 0.15, 0.5 and 1.5 CFUs/mL for the

urine matrix. Swab samples were collected in cobas PCR Media spiked with human colorectal carcinoma HCT-15 cells, and urine samples were collected in upper urinary tract matrix.

Limit of detection analysis of assays using a third-party control diluted in cobas PCR Media or molecular grade water

Amplirun® Total CT/NG/TV/MGE control material (Vircell) was serially diluted using either cobas PCR Media or molecular grade water to six dilution points with nominal concentrations of 111, 36, 8, 4, 1, 0.1 copies/mL in each media, spanning the limit of detection of the Xpert CT/NG test. The panel of 12 samples (six concentrations in each media) were tested in triplicate using cobas CT/NG, PivNG and Xpert CT/NG assays and results compared.

Reference

1. Arifin WN, Yusof UK. Correcting for partial verification bias in diagnostic accuracy studies: A tutorial using R. *Stat Med* 2022;41:1709–27.

Supplementary Table S1. Frequency of available microscopy and culture NG results, with PivNG results stratified by concordant and discordant NG results from cobas CT/NG and Xpert CT/NG

Specimen site	cobas/Xpert NG results	Microscopy and culture NG result	PivNG result	Frequency
Oropharyngeal swab	Concordant negative	Not available	NG negative	38
	Concordant negative	Not available	NG positive	2
	Concordant positive	Negative	NG positive	1
	Concordant positive	Not available	NG positive	72
	Concordant positive	Positive	NG positive	1
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	4
	Discordant positive/negative	Not available	NG negative	26
	Discordant positive/negative	Not available	NG positive	8
	Discordant positive/negative	Positive	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	2
Vaginal swab	Concordant positive	Negative	NG positive	2
	Concordant positive	Not available	NG positive	56
	Concordant positive	Positive	NG positive	2
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	1
	Discordant positive/negative	Not available	NG negative	6
	Discordant positive/negative	Not available	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	31
Male urine sample	Concordant positive	Not available	NG positive	45
	Concordant positive	Positive	NG positive	11
	Discordant positive/negative	Negative	NG negative	1

	Discordant positive/negative	Not available	NG negative	1
	Discordant positive/negative	Not available	NG positive	1
	Missing cobas/invalid Xpert	Not available	NG negative	25
Rectal swab	Concordant positive	Not available	NG positive	73
	Concordant positive	Positive	NG positive	6
	Discordant positive/negative	Negative	NG negative	1
	Discordant positive/negative	Negative	NG positive	1
	Discordant positive/negative	Not available	NG negative	3
	Discordant positive/negative	Not available	NG positive	2
	Missing cobas/invalid Xpert	Not available	NG negative	26
	Missing cobas/invalid Xpert	Not available	NG positive	1

CT, *Chlamydia trachomatis*; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion

Supplementary Table S2. PivNG PPA and NPA with a composite reference standard adjusted for partial verification bias using Begg and Greenes' method

Specimen site	PivNG	Composite reference standard (cobas and Xpert and culture)				PPA (95% CI)	NPA (95% CI)
		Positive	Negative	Undetermined	Total		
Oropharyngeal swab	Positive	75	6	8	89	100% (99.96, 100.00)	91.04% (84.25, 97.83)
	Negative	0	39	28	67		
	Total	75	45	36	156		
Vaginal swab	Positive	60	1	1	62	100% (99.07, 100.00)	100% (99.93, 100.00)
	Negative	0	1	37	38		
	Total	60	2	38	100		
Male urine sample	Positive	56	0	1	57	100% (99.07, 100.00)	100% (99.93, 100.00)
	Negative	0	1	26	27		
	Total	56	1	27	84		
Rectal swab	Positive	79	1	3	83	100% (99.28, 100.00)	96.66% (90.22, 100.00)
	Negative	0	1	29	30		
	Total	79	2	32	113		

CI, confidence interval; NG, *Neisseria gonorrhoeae*; NPA, negative percent agreement; Piv, pilin inversion; PPA, positive percent agreement

Supplementary Table S3. Summary limit of detection estimates for cobas NG-positive samples using PivNG compared with Xpert CT/NG

Matrix	Assay	LOD probit estimate (CFU/mL)	95% CI
Swab	PivNG	0.043	0.026, 0.199
	Xpert NG	0.326	0.162, 1.671
Urine	PivNG	0.018	0.011, 0.121
	Xpert NG	0.451	0.186, 3.136

CFU, colony-forming unit; CI, confidence interval; LOD, limit of detection; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion

Supplementary Table S4. NPA between a specificity panel of non-gonococcal isolates and PivNG-negative results

Expected result	Number of non-gonococcal isolates*	PivNG (+) results	PivNG (-) results	NPA (%)	95% exact CI
Negative	20	0	20	100	83.2, 100.0

*Non-gonococcal *Neisseria* isolates evaluated were: *N. cinerea*, *N. dentrificans*, *N. elongata subsp. elongata*, *N. elongata subsp. niroreducans*, *N. flava*, *N. flavescens*, *N. subflava*, *N. lactamica*, *N. macacae*, *N. meningitidis Serogroup A*, *N. meningitidis Serogroup B*, *N. meningitidis Serogroup C*, *N. meningitidis Serogroup D*, *N. meningitidis Serogroup W135*, *N. meningitidis Serogroup Y*, *N. mucosa*, *N. perflava*, *N. polysaccharea*, *N. sicca* and *N. subflava*

+, positive; -, negative; CI, confidence interval; NG, *Neisseria gonorrhoeae*; NPA, negative percent agreement; Piv, pilin inversion

Supplementary Table S5. Performance of Xpert CT/NG with Vircell control material diluted to six nominal concentrations using cobas PCR Media or molecular grade water

Vircell control diluted in cobas PCR Media					Vircell control diluted in molecular grade water				
NG copies/mL	Xpert NG*		cobas NG	PivNG	NG copies/mL	Xpert NG		cobas NG	PivNG
	NG2 Ct	NG4 Ct	NG Ct	PivNG Ct		NG2 Ct	NG4 Ct	NG Ct	PivNG Ct
111	32.4	32.7	34.22	30.4	111	31.9	31.8	34.14	30.6
111	32.5	32.6	34.28	30.5	111	31.9	31.5	33.84	30.5
111	32.6	32.2	33.84	30.3	111	32.3	32.1	34.75	30.6
36	35	34.8	36.53	32.2	36	34.6	33.5	35.67	32.2
36	33.9	34.1	35.92	31.8	36	34.4	33.7	36.71	32
36	34.4	33.5	36.02	32.4	36	35	34.4	35.27	32.3
8	Not detected	36.1	36.69	35	8	37.4	Not detected	39.08	34.3
8	Not detected	37.7	37.5	34.5	8	35.7	39.4	37.8	34.3
8	37.2	37.2	39.62	34.3	8	Not detected	37.1	38.28	35
4	Not detected	37.4	40.06	35.7	4	36.1	37.9	39.27	35.1
4	37.7	37.4	38.8	36.5	4	Not detected	37.9	41.67	35.2
4	36.7	Not detected	40.7	35.3	4	Not detected	35.9	Not detected	35.5
1	Not detected	Not detected	Not detected	Not detected	1	Not detected	Not detected	39.15	Not detected
1	Not detected	Not detected	39.83	35.8	1	Not detected	Not detected	Not detected	37.4
1	Not detected	Not detected	Not detected	Not detected	1	Not detected	Not detected	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	42	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	Not detected	Not detected	Not detected
0.1	Not detected	Not detected	Not detected	Not detected	0.1	Not detected	Not detected	Not detected	Not detected

*Xpert CT/NG requires both NG targets (designated by Cepheid as NG2 and NG4) to be detected to assign a positive result

CT, *Chlamydia trachomatis*; Ct, cycle threshold; NG, *Neisseria gonorrhoeae*; Piv, pilin inversion